

KRAUJO KREŠĖJIMO ANALIZATORIAI (KOAGULIATORIAI)

Eil. Nr.	Aprašymas
1.	<div style="text-align: center;"> </div> <p>AJ-1326 ANJUE Dviejų kanalų kraujo krešėjimo analizatorius (koagulatorius)</p> <p>Feature :</p> <ul style="list-style-type: none"> ● Advanced scattered light principle and percentage analysis ensure accurate results. ● Low reagent consumption less than 20uL, open reagent. ● Special testing cup location system, adding sample bracket ● Optional PT Derived method for testing fibrinogen, decrease the dosage of reagent ● Advanced optical light matching system ensure precision result. ● 240* 64 large LCD display. ● Memory for 10.000 testing result ● Internal thermal-sensitive printer ● Electronically-linked pipette optional. ● Perform all routine tests such as PT. APTT. TT Fibrinogen and clotting factors ● RS232 interface, PC connecting <p>Specification:</p> <p>Testing channels: 2 channel; 2 different parameters analysis in the same time. Sample pre-warming position: 16 positions reagent pre-warming position: 4 positions Software timer: 2 Test wavelength: 470nm Sample consume: 20uL-40ul Reagent consume: 20uL-40uL</p> <p>Temp control precision: 37±0.3°C</p> <p>Repeatability: Cv < 2% Coincidence indicator: ≤3% Memory: 10000 test result Notation of Result: S, %, PTR, INR, g/L Operation language: English Screen: 240*128 LCD Power supply: 220VAC±15% 50-60Hz Dimensions:395mm(L)×320mm(W)×135mm(H)</p> <p>Temperature:15°C-30°Cwet≤90%</p> <p>Weight:6Kg Usual Test list: Prothrombin PT Activated Partial thromboplastin time APTT Thrombin Time TT Fibrinogen FIB RT, VT, VIII, IX, XI, XII, II, V, VII, X, PS, PCXa etc.</p>

2.



AJ-1328 ANJUE Keturių kanalų kraujo krešėjimo analizatorius (koagulatorius)

Feature :

- Advanced scattered light principle and percentage analysis ensure accurate results.
 - Low reagent consumption less than 20uL, open reagent.
 - Special testing cup location system, adding sample bracket
 - Advanced optical light matching system ensure precision result.
 - Optional PT Derived method for testing fibrinogen, decrease the dosage of reagent
-
- Top quality thermostat, precision $37\pm 0.3^{\circ}\text{C}$

- Dispense precision: $\text{CV} < 2\%$
- Memory for 10.000 testing result
- Internal thermal printer
- 240* 128 large LCD display.
- Electronically-linked pipette optional
- Perform all routine tests such as PT, APTT, TT Fibrinogen and clotting factors.
- RS232 interface, PC connecting

Specification:

Testing channels: 4 channel; 4 different parameters analysis in the same time.

Sample incubation block: 24 positions;

reagent incubation block: 6 positions

Software timer: 4

Test wavelength: 470nm

Memory: 10000 test result

Temp control precision: $37\pm 0.3^{\circ}\text{C}$

Sample consume: 20uL-40ul

Reagent consume: 20uL-40uL

Coincidence indicator: $\leq 3\%$

Notation of Result: S, %, PTR, INR, g/L

Operation language: English

Power supply: 220VAC $\pm 15\%$ 50-60Hz

Dimensions: 395mm(L) \times 320mm(W) \times 135mm(H)

Temperature: 15°C-30°C wet $\leq 90\%$

Weight: 7Kg

Usual Test list:

Prothrombin PT

Activated Partial thromboplastin time APTT

Thrombin Time TT

Fibrinogen FIB

RT, VT, VIII, IX, XI, XII, II, V, VII, X, PS,

3.



BCA-2048B SEEUCO Dviejų kanalų kraujo krešėjimo analizatorius (koagulatorius)

Specifications

Number of Channels 2 , 37°C±0.2°C

Reagent Storage 4

Sample Preparation Positions 6, 3 with timer

Reagent Volume 100µl

Stray ≤1%

Interchannel Error ≤3%

Linearity Error ≤1%

4.



HY150FA POWEAM Vieno kanalo kraujo krešėjimo analizatorius (koagulatorius)

FEATURES

- Channels: 1 channel
- Test Item: TT, PT, APTT, FIB, HEP, LMWH, PC, Ps,
- Coagulation factors II, V, VII, X, VIII, IX, XI, XII
- Test Method: Photoelectric Magnetic beads
- Test speed : ≤300s
- Sample warming up Positions: 8/8/16
- Reagent warming up Positions: 6/6/6
- Linearity Error ≤1%
- Reagents consumption: 50~200µl
- Reagents: Open
- Working environment: 0°C~ 35°C
- Timer error: ≤0.1%
- Sample Measure Time: 4s~180s
- Resolution: 0.1S
- Thermostat system: 37±1°C
- Storage: UP to 6000 test results

5.



WHY40CA POWEAM Dviejų kanalų kraujo krešėjimo analizatorius (koagulatorius)
FEATURES

- Double magnetic route, stable performance , not liable to interference
 - Dot-matrix LCD monitor
 - 100 assay parameters and 1000 test results can be saved
 - Touch keyboard
 - Thermal printer
 - Reagent-open, less consumption ($\leq 200\mu\text{L}$)
 - Clock CMOS chip inserted
 - Standard 232 serial interface
 - Power supply: $220\text{V}\pm 15\%$, 50Hz, 80VA
 - Test method: magnetic ball method
 - sample consumption: $\leq 200\mu\text{L}$
 - Reagent consumption: $50\mu\text{L}\sim 200\mu\text{L}$
 - Temperature: $37^{\circ}\text{C}\pm 0.5^{\circ}\text{C}$
 - Adding sample: by a timing linkage
 - Precision: $\pm 10\%$; $\text{CV} \leq 10\%$
 - Speed: $\leq 300\text{s}$ (depend on setting)
 - Channel: 4/2/1
 - Sample warming up position: 16/8/8
 - Reagent warming up position: 6/6/6
- Test itmes
- Main test:PT, APTT ,TT ,FIB
 - Others: Hep ,ProC ,ProS
 - Coagulation factor II,V,VII,VIII, IX,X, XI,XII, AT-III, AT, RepT, RVVT, RVVCT, RVVRT, LA

6.



WHY20FB POWEAM Dviejų kanalų kraujo krešėjimo analizatorius (koagulatorius)
FEATURES

- Channels: 2 channel
- Test Item: TT, PT, APTT, FIB, HEP, LMWH, PC, Ps
- Coagulation factors II, V,VII, X, VIII, IX, XI, XII
- Test Method: Photoelectric Magnetic beads

- Test speed $\leq 300s$
- Sample warming up Positions: 8/8/16
- Reagent warming up Positions: 6/6/6
- Linearity Error $\leq 1\%$
- Reagents consumption: 50-200 μ l
- Reagents: Open
- Working environment: 0 °C~35 °C
- Timer error: $\leq 0.1\%$
- Sample Measure Time: 4s~180s
- Resolution: 0.1S
- Thermostat system: 37 \pm 1 °C
- Storage: UP to 6000 test results

7.



WHY40CA POWEAM Keturių kanalų kraujo krešėjimo analizatorius (koagulatorius)

FEATURES

- Double magnetic route, stable performance , not liable to interference
 - Dot-matrix LCD monitor
 - 100 assay parameters and 1000 test results can be saved
 - Touch keyboard
 - Thermal printer
 - Reagent-open, less consumption ($\leq 200uL$)
 - Clock CMOS chip inserted
 - Standard 232 serial interface
 - Power supply: 220V \pm 15%, 50Hz, 80VA
 - Test method: magnetic ball method
 - sample consumption: $\leq 200uL$
 - Reagent consumption: 50uL~200uL
 - Temperature: 37°C \pm 0.5°C
 - Adding sample: by a timing linkage
 - Precision: $\pm 10\%$; CV $\leq 10\%$
 - Speed: $\leq 300s$ (depend on setting)
 - Channel: 4/2/1
 - Sample warming up position: 16/8/8
 - Reagent warming up position: 6/6/6
- Test itmes
- Main test:PT, APTT ,TT ,FIB
 - Others: Hep ,ProC ,ProS
 - Coagulation factor II,V,VII,VIII, IX,X, XI,XII, AT-III, AT, RepT, RVVT, RVVCT, RVVRT, LA

8.



WHY40FB POWEAM Keturių kanalų kraujo krešėjimo analizatorius (koagulatorius)

FEATURES

- Channels: 4 channel
- Test Item: TT, PT, APTT, FIB, HEP, LMWH, PC, Ps
- Coagulation factors II, V, VII, X, VIII, IX, XI, XII
- Test Method: Photoelectric Magnetic beads
- Test speed <300s
- Sample warming up Positions: 8/8/16
- Reagent warming up Positions: 6/6/6
- Linearity Error <1%
- Reagents consumption: 50-200ul
- Reagents: Open
- Working environment: 0°C~35°C
- Timer error: <0.1%
- Sample Measure Time: 4s -180s
- Resolution: 0.1S
- Thermostat system: 37C±1°C
- Storage: UP to 6000 test results